


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-V-HV2-P-CEA*

S.N.: *6401*

Date: *JULY 17 1992*



60 dB ATTEN
100 VOLTS/DIV
PRF = 20 KHZ.

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width: *(TO 50ns)*
- c) Rise time:
0 TO 20 NS
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

+15 VOLTS
250 mA (max)

[Signature]