


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-6A-C-P-ET-EN-LIB*

S.N.: *6405*

Date: *SEPT 9 1992*



- a) Output signal amplitude:  
*0 TO +1.0 AMP*
- b) Pulse width:  
*50 NS TO 5.0 US*
- c) Rise time:  
*≤ 10 NS*
- d) Fall time:  
*≤ 10 NS*
- e) PRF:  
*0 TO 1.0 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power: *120 / 240V*  
*50-60 Hz*

ⓐ *MANFOLDING OUT TO RL = 50 Ω (200 MHz SCOPE)*

