


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-3A-COA

S.N.: 6409

Date: AUG 7 1992



60 dB ATTEN

100 VOLTS/DIV

PRF = 10 KHz

a) Output signal amplitude:  
+150 TO +450 V

b) Pulse width:  
TO 50 ns  
≤ 4 ns

c) Rise time:  
≤ 1.0 ns

d) Fall time:  
≤ 2.5 ns

e) PRF:  
0 TO 20 KHz

f) Jitter, stability:  
OK

g) Prime power:  
+15 VOLTS  
400 mA (MAX)