


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AV-143 - IBAMC

S.N.: 6414

Date: SEPT 2 1992



- a) Output signal amplitude:  
0 TO +30 VOLTS TO
- b) Pulse width:  $R_L > 1K$   
100 NS TO DC
- c) Rise time:  $\leq 60 NS$
- d) Fall time:  $\leq 60 NS$
- e) GAIN:  $\times 6$  (VOLTAGE)
- f) Jitter, stability: OK
- g) Prime power:

$R_L = 1K.$

+24 V  
450 mA. (max)

h)  $R_{in} : 50 \Omega$