


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVR-5B-17-PS-602-M-PWT-AT-R4*

S.N.: *6423*

Date: *AUG 24 1992*



*R_L = 50 Ω
100 VOLT/DIV
PRF = 500 Hz*



*R_L = 50 Ω
100 VOLT/DIV
PRF = 50 Hz*

- a) Output signal amplitude:
0 TO +500 VOLTS TO
- b) Pulse width: *R_L > 50 Ω
0.1 TO 100 μ S*
- c) Rise time: *(1% MAX DUTY CYCLE)
≤ 30 NS*
- d) Fall time:
≤ 30 NS
- e) PRF: *0 TO 10 KHz
(1% MAX DUTY)*
- f) Jitter, stability: *(CYCLES)
OK*
- g) Prime power:
*120/240 VOLTS
50-60 Hz.*

[Signature]