


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-5-HVI-N

S.N.: 6426

Date: OCT 1 1992



- a) Output signal amplitude:
0 to -40 VOLTS
- b) Pulse width:
± 350 ps (FWHM)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:

± 15 VOLTS,
175mA (MAX)

40 db ATTEN
10 VOLTS/DIV
PRF = 10 MHz

[Handwritten signature]