


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-3 PW-PS-P-USA*

S.N.: *6437*

Date: *AUG 20 1992*



R_L = 10K
1/f = 100 Hz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width: *(R_L > 5K)*
0.5 TO 50 nS
- c) Rise time:
≤ 20 nS
- d) Fall time:
≤ 20 nS
- e) PRF:
MIN PW: 400 Hz (20% MAX)
MAX PW: 4 Hz (OUTY CYCLE)
- f) Jitter, stability:
- g) Prime power: *OK*
120/200 VOLTS,
50-60 Hz