


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-3A-COB*

S.N.: *6439*

Date: *OCT 19 1992*



A ↑ ↑ B
60 db ATTEN
100 VOLTS/DIV
PRF = 10 KHZ

- a) Output signal amplitude:
+200 TO +450 VOLTS
- b) Pulse width: *8 NS*
- c) Rise time: *~ 3 NS*
- d) Fall time: *~ 3 NS*
- e) PRF: *0 TO 20 KHZ*
- f) Jitter, stability:
OK
- g) Prime power:

+15 VOLTS
700 mA (MAX)
WA