


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AV-1-P*

S.N.: *6448*

Date: *SEPT 11 1992*



- a) Output signal amplitude:  
*0 TO + 10 VOLTS*
- b) Pulse width:  
*0.2 TO 40 NS*
- c) Rise time:  
*≤ 100 PS*
- d) Fall time:  
*≤ 135 PS*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*40 dB ATTEN*  
*10 VOLTS DIV*  
*PRF = 300 KHz*

*+15 VOLTS*  
*250 mA (MAX)*

*[Signature]*