


PULSE GENERATOR
PERFORMANCE CHECK


Model: AVR-5A-C-PWT-P

S.N.: 6456

Date: SEPT 10 1992



Ⓐ 100 VOLTS/DIV
 $R_L = 50 \Omega$
PRF = 2.0 KHz



Ⓑ 100 VOLTS/DIV
 $R_L = 50 \Omega$
PRF = 2.0 KHz.

- a) Output signal amplitude:
0 TO +500 VOLTS
- b) Pulse width: TO $R_L \geq 50 \Omega$
0.1 TO 5.0 μ S
- c) Rise time:
 ≤ 30 NS
- d) Fall time:
 ≤ 30 NS
- e) PRF: 0 TO 10 KHz
(1% MAX DUTY)
- f) Jitter, stability: (CYCLE)
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz