


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVRH-2-C-P-PWT-AT

S.N.: 6460

Date: SEPT 29 1992



- a) Output signal amplitude:
0 TO +2000 VOLTS
- b) Pulse width: $R_L \geq 1000 \Omega$
200 NS TO 5.0 μ S
- c) Rise time: ≤ 60 ns
- d) Fall time: ≤ 60 ns
- e) PRF: 1 Hz TO 1 KHz
(OR SINGLE PULSES)
- f) Jitter, stability: OK
- g) Prime power: 120 / 240 VOLTS
50 - 60 Hz.

HORIZ: 500 ns/DIV
VERT: 500 VOLTS/DIV
PRF: 1 KHz

