


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-3A-COA1*
S.N.: *6469*
Date: *OCT 7 1992*



- a) Output signal amplitude:
150 TO 450 VOLTS
- b) Pulse width:
7 NS (AT 20% POINT)
- c) Rise time:
≈ 3 NS
- d) Fall time:
≈ 3 NS
- e) PRF:
0 TO 20 KHz.
- f) Jitter, stability:
OK
- g) Prime power:

60. db ATTEN
100 VOLTS/DIV
PRF ≈ 10 KHz.

+15 VOLTS
350 mA (MAX)