


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PRR-A-1-PW-PS-05-N*

S.N.: *6471*

Date: *OCT 2 1992*



$$R_c = 50 \Omega$$

$$PRF = 10 \text{ kHz}$$

- a) Output signal amplitude:
0 TO - 200 VOLTS
- b) Pulse width:
50 NS TO 1.0 US
(0.5% MAX DUTY CYCLE)
- c) Rise time:
≤ 60 NS
- d) Fall time:
≤ 10 NS
- e) PRF: *0 TO 100 kHz*
(0.5% MAX DUTY CYCLE)
- f) Jitter, stability: *OK*
- g) Prime power:
120/240V
50-60 Hz