


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-3-PN-SRIA-PS

S.N.: 6473

Date: OCT 9 1992



- a) Output signal amplitude:
0 TO ± 200 VOLTS
- b) Pulse width: ($R_L > 50 \Omega$)
1 μ S TO 10 MS
- c) Rise time: (50% MAX DUTY CYCLE)
 ≤ 30 NS
- d) Fall time:
 ≤ 30 NS
- e) PRF: DC TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

Point $R_L = 10K$.

$V_{DC IN} = +200$ VOLTS

a) 120/240 V 50-60 Hz

b) 0 TO ± 200 VOLTS
DC (200 mA MAX)