


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVN-2-P*

S.N.: *6474*

Date: *NOV 3 1992*



- a) Output signal amplitude:
0 TO +2 (USEFUL TO 10 VOLTS)
- b) Pulse width:
0.3 TO 1.5 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 100 PS
- e) PRF: *50 TO 250 MHz*
- f) Jitter, stability:
O/K
- g) Prime power:

*40 dB ATTEN
10 VOLTS/DIV
PRF = 130 MHz*

*+24 VOLTS
450 mA (MAX)*

[Signature]