


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MVP-MV-1-P*

S.N.: *6483*

Date: *OCT 30 1992*



- a) Output signal amplitude:
0 TO 710 VOLT
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 dB ATTEN
10 VOLT/DIV
300 KHz PRF

715 VOLT
200 mA (max)
[Signature]