


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-1-7*
S.N.: *6484*
Date: *OCT 30 1992*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
0.2 TO 4.0 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 100 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
200mA (max)

40 dB ATTEN
10 VOLTS/DIV
300 MHz PRF