


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PKR-17*

S.N.: *6485*

Date: *OCT 30 1992*



40.26 ATTEN
5 VOLTS/DIV
300 KHZ PRF

- a) Output signal amplitude:
≥ +5 VOLTS
- b) Pulse width:
40 NS TO 100 NS
- c) Rise time:
SEE 1716 (42)
≤ 100 PS
- d) Fall time:
≤ 2 NS
- e) PRF: *0 TO 1.0 MHz (MAX PW)*
0 TO 20 KHz (MAX PW)
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
200 mA (MAX)

[Signature]