


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANMP-4-C-P*


S.N.: *6490*

Date: *NOV 20 1992*



- a) Output signal amplitude: *0 TO +10 VOLTS TO*
- b) Pulse width: *50 ns*
10 NS TO 10 US
(10% MAX DUTY CYCLE)
- c) Rise time: *≤ 200 PS*
- d) Fall time: *≤ 200 PS*
- e) PRF: *0 TO 100 MHz*
(10% MAX DUTY CYCLE)
- f) Jitter, stability: *OK*
- g) Prime power: *120/200 VOLTS*
50-60 Hz

R_L = 50 Ω, PRF = 1.0 KHz



← 5ns/div

← 500 ns/div
(PULSE TIME)

40 dB ATTEN PRF = 100 KHz

10 VOLTS/DIV