


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-MV-C-P*

S.N.: *6494*

Date: *NOV 18 1992*



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
1-50 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 2 NS
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability:
OK

40 db ATTEN

10 VOLTS/DIV

TOP: 5 NS/DIV

BOT: 200 PS/DIV

(RISE TIME)

- g) Prime power:
120/240 V
50-60 Hz.
- [Signature]*