


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-5A-C-P*
S.N.: *6500*
Date: *NOV 24 1992*



- a) Output signal amplitude: *0 TO + 40 AMPS*
- b) Pulse width: *TO 0.7 μ*
20 TO 150 ns
- c) Rise time: *$\leq 5 ns$*
- d) Fall time: *$\leq 5 ns$*
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability: *OK*

$R_L = 0.7 \Omega$
 $\approx 14 AMPS / DIV$
PRF = 5.0 KHz

g) Prime power: *120 / 240 V*
50 - 60 Hz
[Signature]