


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-3A-P-C-LIA*

S.N.: *6510*

Date: *DEC 2 1992*



- a) Output signal amplitude:
+2 TO +75 AMP
- b) Pulse width:
2.5 NS (HALF POWER)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

$$R_{SI} = 5.1 \Omega$$

$$R_{SC} = 10 \Omega$$

$$\therefore \approx 20 \text{ AMPS/DIV}$$

$$120/240 \text{ VOLTS}$$

$$50 \text{ } \approx \text{ } 60 \text{ MHz}$$