


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-S-1-N-C*

S.N.: *6511*

Date: *DEC 3 1992*



*40 dB ATTEN*

*10 VOLTS/DIV*

*PRF = 1.0 MHz*

a) Output signal amplitude:

*0 TO -10 V*

b) Pulse width:

*≤ 130 ps*

c) Rise time:

*NA*

d) Fall time:

*NA*

e) PRF:

*0 TO 1.0 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*120 / 240 VOLTS  
50-60 Hz*