


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-1N-15-C-P*

S.N.: *6516*

Date: *NOV 24 1992*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 60 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/200 V
50/60 MHz

[Signature]

40 dB ATTEN
5 VOLTS/DIV
TOP 1 NS/DIV
BOT 100 PS/DIV (RISE TIME)

PRF = 300 KHz