


PULSE GENERATOR

PERFORMANCE CHECK

Model: PVH-5-1-P-PS

S.N.: 6517

Date: NOV 24 1992



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
≤ 130 PS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

40 dB ATTEN

10 VOLTS/DIV

PRF = 300 KHz