


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVG-3A-COAI*

S.N.: *6521*

Date: *DEC 10/92*



60 db ATTEN
100 VOLTS/DIV
PRF \approx 10 KHz

- a) Output signal amplitude:
+ 150 TO +450 VOLTS
- b) Pulse width:
7 NS (AT 20% POINT)
- c) Rise time:
 \approx 3 NS
- d) Fall time:
 \approx 3 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+ 15 VOLTS
350 mA (MAX)
- h) Spurious:
As per COBI units