


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-6A-C-P-CTR1*


S.N.: *6534*

Date: *JAN 13 1992*



Ⓐ MAIN FRAME OUT
TO 50-Ω
(50 dB ATTEN)
∴ 32 VOLTS/DIV
PRF = 10 KHz

- a) Output signal amplitude:
0 TO +1.0 Amps
- b) Pulse width:
20 TO 200 NS
- c) Rise time:
≤ 1 NS
- d) Fall time:
≤ 1 NS
- e) PRF:
0 TO 10 KHz
- f) Jitter, stability:
OK



- g) Prime power:
120/240 VOLTS
50-60 Hz

Ⓑ ME OUTPUT OF
AVX-51-ME-MV-CTR1
MODULE (30 dB ATTEN)