


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV6-3A-COA1*

S.N.: *6542*

Date: *JAN 15 1993*



- a) Output signal amplitude:
+150 TO 450 VOLTS
- b) Pulse width:
7 NS (20% RISE POINT)
- c) Rise time:
≈ 3 NS
- d) Fall time:
≈ 3 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

60 dB ATTEN
100 VOLTS/DIV
PRF ≈ 10 KHz

+15 VOLTS
550 mA (max)
h) SPURIOUS
NS PER
COB / UNIT?
