


PULSE GENERATOR

PERFORMANCE CHECK

Model: *M6-3A. -CoB1*

S.N.: *6543*


Date: *JAN 15 1993*



- a) Output signal amplitude:
+200 TO +450 V
- b) Pulse width:
≈ 8 NS
- c) Rise time:
≈ 3 NS
- d) Fall time:
≈ 3 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*+15 VOLTS
900 mA MAX*

[Signature]



*60 dB ATTEN
- 100 VOLTS/DIV
PRF ≈ 10 KHz*