


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-2A-P

S.N.: 6544

Date: JAN 20 '93



60 dB ATTEN

100 VOLTS/DIV

PRF = 5.0 KHz

- a) Output signal amplitude:  
0 TO +170 V (TO)
- b) Pulse width: 50ns
- c) Rise time: 3 TO 100 NS  
≤ 2 NS
- d) Fall time: < 2 NS
- e) PRF: 0 TO 5 KHz
- f) Jitter, stability: OK
- g) Prime power:

+15 VOLTS  
500 mA (MAX)