


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PUL-2A-P*

S.N.: *6545*

Date: *JAN 20 '93*



*60 db ATTEN*

*100 VOLTS/DIV*

*PRF = 5.0 KHz*

a) Output signal amplitude:

*0 TO +100 VOLTS*

b) Pulse width: *CTD 50 ns*

c) Rise time: *3 TO 100 ns*

*≤ 2 ns*

d) Fall time:

*≤ 2 ns*

e) PRF:

*0 TO 5 KHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+ 15 VOLTS*

*500 mA (max)*