


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PAH-5-1-C-P*

S.N.: *6547*

Date: *JAN 28 1993*



a) Output signal amplitude:
0 TO +10 VOLTS

b) Pulse width:
≤ 130 PS

c) Rise time:
NA


d) Fall time:
NA

e) PRF:
0 TO 1-6 MHz

f) Jitter, stability:
OK

g) Prime power:
120/240 VOLTS
50-60 Hz

*ⓐ LED db RETURN
10 VOLTS/DIV
PRF = 100 KHz*



*ⓑ AS ⓐ BUT
500 PS/DIV.*