


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PM-2A-C-N*

S.N.: *6551*

Date: *JAN 28 1993*



- a) Output signal amplitude:  
*0 to -170 V<sub>OUT</sub>*
- b) Pulse width: *50 ns*
- c) Rise time:  
*3 to 100 ns*  
*≤ 2 ns*
- d) Fall time:  
*≤ 2 ns*
- e) PRF:  
*0 to 5 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*60 db ATTEN*  
*100 V<sub>OUT</sub>/DIV*  
*PRF = 5.0 kHz*

*120/240 V*  
*50 to 100 Hz*  
*A*