


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVRL-1-PS-05*

S.N.: *6560*

Date: *FEB 3 1993*



*60 db ATTEN*

*100 VOLTS/DIV*

*PRF = 3.0 KHZ.*

- a) Output signal amplitude:  
*-200 VOLTS (7050M)*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 3 NS*
- e) PRF:  
*0 TO 5 KHZ.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 VOLTS*  
*50 ~ 60 Hz*
- h) OS: *0 TO ± 50 VOLTS*  
*(1 mA MAX)*

*AB*