


PULSE GENERATOR

PERFORMANCE CHECK


Model: *MR-61-PN-OT-C-T1B*

S.N.: *6575*

Date: *MAR 4 1993*



*(A) Pout WITH -50V OFFSET.  
PRF = 1.0 MHz*



*(B) AS (A) BUT 50NS/DIV*

- a) Output signal amplitude: *0 TO 250 VOLTS*
- b) Pulse width: *TO R<sub>L</sub> > 10K  
10 NS TO 1.0 US*
- c) Rise time: *≤ 7 NS*
- d) Fall time: *≤ 7 NS*
- e) PRF: *0 TO 1.0 MHz  
(10% MAX DUTY)*
- f) Jitter, stability: *CYCLE  
OK*
- g) Prime power: *120/240 V  
50-60 Hz.*