


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-A-1-C-P-PW-M*

S.N.: *6577*

Date: *FEB 25 1993*



40 dB ATTEN

10 VOLTS DIV

PRF = 300 kHz.

- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width: *(TO 50ns)*
0.2 TO 40 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz
[Signature]