


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MU-LDD2-C-P*

S.N.: *6578*

Date: *APRIL 7 1993*



- a) Output signal amplitude:
0 TO +40 AMP
- b) Pulse width: (*TO $R_L = 5.0 \Omega$*)
20 NS TO 1.0 μ S
- c) Rise time: (*0.06% MAX DUTY CYCLE*)
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF: *0 TO 5 KHz*
(0.06% MAX DUTY)
- f) Jitter, stability: *checked*
- g) Prime power: *OK*

$R_L = 5.0 \Omega$

PRF = 2.0 KHz

220 V 50-60 Hz