


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-1-P-LT2B*

S.N.: *6579*

Date: *MAR 18 1993*



60 dB ATTEN
100 VOLTS/DIV
PRF = 10 KHz

- a) Output signal amplitude:
≥ 250 VOLTS
- b) Pulse width: *TO 50ns*
≤ 4 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS DC
150 mA MAX

[Signature]