


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV14-MV1-P-LT2B

S.N.: 6580

Date: MAR 18 1993



60 db ATTEN

100 VOLTS/DIV

PRF = 100 KHz

- a) Output signal amplitude:
7.1 + 100 VOLTS
- b) Pulse width: (PD 50-n)
≤ 1.0 NS
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 100 KHz
- f) Jitter, stability: OK
- g) Prime power: +15 VDC
125 mA MAX

