


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ARRH-1-C-05-P*

S.N.: *6583*

Date: *APRIL 12 1993*



- a) Output signal amplitude:
0 TO +1000 VOLTS
- b) Pulse width: $R_L \geq 200 \Omega$
200ns to 5.0us
- c) Rise time:
 $\leq 50 ns$
- d) Fall time:
 $\leq 50 ns$
- e) PRF: *0 TO 1 KHz*
- f) Jitter, stability:
OK

200 VOLTS/DIV
PRF = 1.0 KHz
05 = +200 VOLTS
 $R_L = 10 K$

- g) Prime power:
120/240 Volts,
50 x 60 Hz
- h) 05: *0 to $\pm 200 volts,$*
0 to $\pm 50 mA.$

MA