


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-1-P-LT2B

S.N.: 6598

Date: MAR 18 1993



60 dB ATTEN  
100 VOLTS/DIV  
10. KHZ PRF.

- a) Output signal amplitude:  
 $\geq +250$  VOLTS
- b) Pulse width: TO 50n  
 $\leq 4$  NS
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF: 0 TO 50 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 VDC  
150 mA (MAX)