


PULSE GENERATOR

PERFORMANCE CHECK

Model: *M-140X-CA2*

S.N.: *6603*

Date: *MAY 28 1993*



a) Output signal amplitude:  
*+600V TO RL > 50Ω*

b) Pulse width:  
*2.0 μS.*

c) Rise time:  
*≤ 30 ns*


d) Fall time:  
*≤ 30 ns*

e) PRF:  
*0 TO 100 Hz*

f) Jitter, stability:  
*OK*

g) Prime power:  
*+12 VDC*  
*(+10 TO +16 VDC)*  
*2.0 AMP (MAX)*

*≈ 100 VOLTS / DIV*



*[Signature]*