


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PM-5-1-P-BLC*

S.N.: *6607*

Date: *APRIL 19 1993*



- a) Output signal amplitude:
7.12 VOLTS
- b) Pulse width:
70 ps (FWHM)
ADJUSTABLE FROM 0 TO 1.0 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

40 dB/DIV
10 VOLTS/DIV
PRF = 300 kHz
TOP 100 ps/DIV
BOOT 1.0 ns/DIV

15 VOLTS
200 mA MAX
[Signature]