


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-V-DR-C-P-URA*

S.N.: *4638*

Date:



- a) Output signal amplitude:
> 35 VOLTS
- b) Pulse width: *(to 50%)*
2 TO 20 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 2 NS
- e) PRF: *0 TO 20 kHz*
- f) Jitter, stability:
OK
- g) Prime power:

TDP 50 db
32 VOLTS/DIV
5 NS/DIV

120/240 V
50-60 Hz

BOT: 100 PS/DIV
(RISE TIME)