


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-2A-C-N-M*

S.N.: *6639*

Date: *APRIL 1 1993*



- a) Output signal amplitude:  
*0 TO 5 Amp*
- b) Pulse width: *(20 to 200 ns)*  
*2 TO 50 ns*
- c) Rise time:  
*≤ 1 ns*
- d) Fall time:  
*≤ 2 ns*
- e) PRF:  
*0 TO 20 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*≈ 3 Amp / 10 Ω*  
*R<sub>L</sub> = 10 Ω*  
*PRF = 20 kHz*

*120 / 240 V*  
*50 - 60 Hz*  
*MJ*