


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVG-1-P-LT2B*

S.N.: *6641*

Date: *APR 6/93*



- a) Output signal amplitude:
 $\geq +250$ VOLTS TO 50 Ω
- b) Pulse width:
 ≤ 4 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
+ 15 VDC
200 mA (MAX)

60 db ATTEN

100 VOLTS/DIV

PRF = 10 KHZ