


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVG-1-P-LT2B*

S.N.: *6642*

Date: *APR 6 193*



- a) Output signal amplitude:  
*7 + 250 VOLTS TO 50Ω*
- b) Pulse width:  
*≤ 4 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 50 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+ 15 VDC*  
*200 mA (MAX)*

*60 db ATTEN*  
*∴ 100 VOLTS/DIV*  
*PRF = 10 KHZ*