


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-HV1-P-LT2B

S.N.: 6643

Date: APR 6/93



- a) Output signal amplitude:  
 $\geq + 100 \text{ VOLTS (TO } 50 \Omega)$
- b) Pulse width:  
 $\leq 1.0 \text{ NS}$
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 100 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:

60 db ATTEN

$\therefore 100 \text{ VOLTS / DIV}$

PRF = 100 KHZ

+ 15 VDC

150 mA (MAX)

*Handwritten signature*