


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-HVI-P-LT2B

S.N.: 6644

Date: APR 6/93



- a) Output signal amplitude:  
 $\geq + 100 \text{ VOLTS (TO } 50 \Omega)$
- b) Pulse width:  
 $\leq 1.0 \text{ NS}$
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 100 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
+ 15 VDC  
150 mA (MAX)

60 db ATTEN  
∴ 100 VOLTS/DIV  
PRF = 100 KHz