


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-HVI-P-LT2B

S.N.: 6645

Date: APR 6/93



- a) Output signal amplitude:
 $\geq + 100 \text{ VOLTS (TO } 50 \Omega)$
- b) Pulse width:
 $\leq 1.0 \text{ NS}$
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+ 15 VDC
150 mA (MAX)

60 db ATTEN
100 VOLTS/DIV
PRF = 100 KHz